Serial No. C 133095

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com

WEIGHTRON BILANCIAI
INDUSTRIAL WEIGHING SPECIALISTS
BILANCIAI GROUP

Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com

UKWF
Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

EDWARDS RESTAURANT LTD
PERRY RD
DAGENHAM
LOCATION RM9 6QD

Machine Details

Cust ref	~~MAIN~~ IN
Make	BILANCIAI
Model	~~DP500~~ FLYNET
Ser No	2503 7142
Capacity	50000 / 1kg /
e/d	20 / 1kg /
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
WBSD-71	TRI30711	3-10-25

Cert Interval	1 YEAR	
Next Cert	MARCH	2027

Tolerance +/-	1kg
0- 10000	20
10000-40000	40
40000-50000	60

Comments ... + MIXED D WEIGHTS
FLT = 5300
TEST UNITS 20580

2 4 6
1 3 5
TX

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
2000	2000	0	2000	0
4000	4000	0	4000	0
8000	8000	0	8000	0
16000	16000	0	16000	0
23300	23300	0	23300	0
30580	30580	0	30580	0
43880	43920	+40	43920	+40

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
43880	43920	43920	43920

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 9300 kg

Reading		Dev	Reading		Dev
1	9360	0	2	9320	+20
3	9300	0	4	9300	0
5	9300	0	6	9300	0
7			8		

Eccentric Load

Reading		Dev	Reading		Dev
1			2		
3			4		
5			6		
7			8		

Signature ..... Name N BOTTEN Date 19.3.26

C1.3.88.1N