

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

EDWARDS WASTE PAPER
GALLIANS CLOSE
BARKING
ESSEX
LOCATION

Machine Details

Cust ref	
Make	BILANCIAI
Model	FLYNET 50
Ser No	T211702432
Capacity	5000 kg
e/d	20kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
W550712211220		2/4/23

Cert Interval	12 MONTHS
Next Cert	MAR 2025

Tolerance +/-	
0-10000	+/- 20
1000-10000	+/- 40
1000-50000	+/- 60

Comments


MAKE UP WEIGHTS USED

2	4	6
1	3	5
		x

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
2000	2000	0	2000	0
10000	10000	0	10000	0
18000	18000	0	18000	0
24140	24140	0	24140	0
34140	34140	0	34140	0
43780	43780	0		


DEFINITIVE LINEARITY & HYSTERESIS



Repeatability

Load	Ind	Ind	Ind
43780	43780	43780	43780


Repeatability



Eccentric Load

Reading	Dev	Reading	Dev
1 9640	0	2 9640	0
3 9640	0	4 9620	-20
5 9640	0	6 9640	0
7		8	

Eccentric Load



Signature [Signature] Name K. LAMPARD Date 26/3/24
C1.3.88.1N