

**Weightron Bilanciai Ltd,**  
Titan Works  
Bridge Way off Broombank Road,  
Chesterfield Trading Estate,  
Chesterfield S41 9QJ, UK  
Tel: 01246 260062  
Fax: 01246 260844  
www.weightron.com



**Weightron Bilanciai Ltd,**  
Unit 4b, Block 4  
Westpark, Chelston  
Wellington TA21 9AD, UK  
Tel: 01823 662355  
Fax: 01823 662017  
www.weightron.com



## CALIBRATION CERTIFICATE

*This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL*

**Customer**

Edwards Recycling Thames Road Barking Essex IG11 0SD LOCATION
--

**Machine Details**

Cust ref	Weybridge
Make	Bilanciai
Model	FLYnet 50
Ser No	T211702432
Capacity	50000 kg
e/d	20kg / 20kg
Working cap	50000kg (400kg)

**Test Weights**

Set No.	Cert No.	Date Cal
PDA-22	TR11747	2/12/22

Cert Interval	12M
Next Cert	March 2024

Tolerance +/-	
0 - 10000	20
10000 - 40000	40
40000 - 50000	60


**Comments** All Readings are in kg  
Weybridge Calibrated and within  
tolerance at time of test.

Make up Weight used on this  
Calibration

2	4	6
1	3	5


**AS FOUND LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
6000	6020	20	6020	20
10000	10020	20	10020	20
20000	20020	20	20020	20
25540	25540	0	25540	0
38700	38680	20	38680	20
44240	44220	20		




**Repeatability**

Load	Ind	Ind	Ind
44240	44220	44220	44180



**Eccentric Load 10000kg**

Reading	Dev	Reading	Dev
1 10020	20	2 10020	20
3 10000	0	4 10000	0
5 9980	20	6 9980	20
7		8	



Signature ..... *M. Rose* Name M. Rose Date 28/3/23

C1.3.88.1N